



Kansas City

An Independent Licensee of the Blue Cross and Blue Shield Association

Postsurgical Outpatient Use of Limb Compression Devices for Venous Thromboembolism Prophylaxis

Policy Number: 1.01.28

Origination: 4/2013

Last Review: 4/2014

Next Review: 4/2015

Policy

Blue Cross and Blue Shield of Kansas City (Blue KC) will provide coverage for outpatient use of limb compression devices when it is determined to be medically necessary because the criteria shown below are met.

When Policy Topic is covered

Outpatient use of limb compression devices for venous thromboembolism prophylaxis after major orthopedic surgery may be considered **medically necessary** in patients with a contraindication to pharmacological agents i.e., at high-risk for bleeding.

Outpatient use of limb compression devices for venous thromboembolism prophylaxis after major non-orthopedic surgery may be considered **medically necessary** in patients who are at moderate or high risk of venous thromboembolism (see Considerations) with a contraindication to pharmacological agents i.e., at high-risk for bleeding.

When Policy Topic is not covered

Outpatient use of limb compression devices for venous thromboembolism prophylaxis after major orthopedic surgery is considered **investigational** in patients without a contraindication to pharmacological prophylaxis.

Outpatient use of limb compression devices for venous thromboembolism prophylaxis after major non-orthopedic surgery or nonmajor orthopedic surgery is considered **investigational** in patients who are at moderate or high risk of venous thromboembolism without a contraindication to pharmacological prophylaxis and in patients who are at low-risk of venous thromboembolism.

Outpatient use of limb compression devices for venous thromboembolism prophylaxis after all other surgeries is considered **investigational**.

Outpatient use of limb compression devices for venous thromboembolism prophylaxis for periods longer than 30 days post-surgery is **not medically necessary**.

Considerations

For purposes of this policy, "major orthopedic surgery" includes total hip arthroplasty, total knee arthroplasty, or hip fracture surgery.

Guidance on determining high risk for bleeding

The ACCP guidelines on prevention of VTE in orthopedic surgery patients list the following general risk factors for bleeding (3):

- Previous major bleeding (and previous bleeding risk similar to current risk)

- Severe renal failure
- Concomitant antiplatelet agent
- Surgical factors: history of or difficult-to-control surgical bleeding during the current operative procedure, extensive surgical dissection, and revision surgery

The guidelines note, however, that “specific thresholds for using mechanical compression devices or no prophylaxis instead of anticoagulant thromboprophylaxis have not been established.”

A clinical guideline from the American Academy of Orthopaedic Surgeons (2011) states (6):

"Patients undergoing elective hip or knee arthroplasty are at risk for bleeding and bleeding-associated complications. In the absence of reliable evidence, it is the opinion of this work group that patients be assessed for known bleeding disorders like hemophilia and for the presence of active liver disease which further increase the risk for bleeding and bleeding-associated complications. (Grade of Recommendation: Consensus) Current evidence is not clear about whether factors other than the presence of a known bleeding disorder or active liver disease increase the chance of bleeding in these patients and, therefore, the work group is unable to recommend for or against using them to assess a patient's risk of bleeding. (Grade of Recommendation: Inconclusive)"

Guidance on duration of use

In patients with contraindications to pharmacologic prophylaxis who are undergoing major orthopedic surgery (THA, TKA or HFS), the ACCP guidelines are consistent with use of intermittent compression devices for 10-14 days after surgery. (3) The ACCP suggestion on extended prophylaxis (up to 35 days) was a weak recommendation that did not mention compression devices as an option.

In the ACCP guideline on VTE prophylaxis in patients undergoing non-orthopedic surgery, the length of standard duration or “limited duration” prophylaxis was not defined. However, “extended duration” pharmacologic prophylaxis was defined as 4 weeks; this was recommended only for patients at high risk for VTE undergoing abdominal or pelvic surgery for cancer and not otherwise at high risk for major bleeding complications.

Guidance on risk level for patients undergoing non-orthopedic surgery

The ACCP guidelines on prevention of VTE in non-orthopedic surgery patients included the following discussion of risk levels (5): (pp. 13-14)

“In patients undergoing general and abdominal-pelvic surgery, the risk of VTE varies depending on both patient-specific and procedure-specific factors. Examples of relatively low-risk procedures include laparoscopic cholecystectomy, appendectomy, transurethral prostatectomy, inguinal herniorrhaphy, and unilateral or bilateral mastectomy. Open abdominal and open-pelvic procedures are associated with a higher risk of VTE. VTE risk appears to be highest for patients undergoing abdominal or pelvic surgery for cancer... Patient-specific factors also determine the risk of VTE, as demonstrated in several relatively large studies of VTE in mixed surgical populations. Independent risk factors in these studies include age at least 60 years, prior VTE, and cancer; age >60 years, prior VTE, anesthesia at least 2 h, and bed rest at least 4 days; older age, male sex, longer length of hospital stay, and higher Charlson comorbidity score; and sepsis, pregnancy or postpartum state, central venous access, malignancy, prior VTE, and inpatient hospital stay more than 2 days. In another study, most of the moderate to strong independent risk factors for VTE were surgical complications, including urinary tract infection, acute renal insufficiency, postoperative transfusion, perioperative myocardial infarction, and pneumonia.“

The American College of Obstetricians and Gynecologists (ACOG) proposed the following risk classification for VTE in patients undergoing major gynecological surgery (available online at: <http://guidelines.gov/content.aspx?id=11429>):

Low: Surgery lasting less than 30 minutes in patients younger than 40 years with no additional risk factors.

Moderate: Surgery lasting less than 30 minutes in patients with additional risk factors; surgery lasting less than 30 minutes in patients age 40-60 years with no additional risk factors; major surgery in patients younger than 40 years with no additional risk factors.

High: Surgery lasting less than 30 minutes in patients older than 60 years or with additional risk factors; major surgery in patients older than 40 years or with additional risk factors.

Highest: Major surgery in patients older than 60 years plus prior venous thromboembolism, cancer, or molecular hypercoagulable state.

Description of Procedure or Service

Patients undergoing major orthopedic surgery are at increased risk for venous thromboembolism (VTE). Patients undergoing other types of surgery may also be at increased risk of VTE. Limb compression devices are one option for thromboprophylaxis and are commonly used in the hospital setting. Outpatient use of compression devices following hospitalization, with or without pharmacologic prophylaxis, has also been proposed.

Background

Patients undergoing major surgery are at increased risk of developing deep vein thrombosis (DVT) and pulmonary embolism (PE), together known as venous thromboembolism (VTE). Patients who are having major orthopedic surgery (defined here as total hip arthroplasty [THA], total knee arthroplasty [TKA] and hip fracture surgery [HFS]) are at particularly high risk. Risk of DVT is increased due to venous stasis of the lower limbs as a consequence of immobility during and after surgery. In addition, direct venous wall damage associated with the surgical procedure itself may occur. DVTs are frequently asymptomatic and generally resolve when mobility is restored. However, some episodes of acute DVT can be associated with substantial morbidity and mortality. The most serious adverse consequence of an acute DVT is a PE which can be fatal; this occurs when the DVT detaches and migrates to the lungs. In addition, DVT may produce long-term vascular damage that leads to chronic venous insufficiency. Without thromboprophylaxis, the incidence of venographically detected DVT is approximately 42-57% after total hip replacement, and the risk of pulmonary embolism is approximately 1-28%. (1) Other surgical patients may also be at increased risk of VTE during and after hospitalization. For example, it is estimated that rates of VTE without prophylaxis after gynecologic surgery is about 15-40%. (2)

Thus, antithrombotic prophylaxis is recommended for patients undergoing major orthopedic surgery and other surgical patients at increased risk of VTE. For patients undergoing major orthopedic surgery, clinical practice guidelines published in 2012 by the American College of Chest Physicians (ACCP) recommend that one of several pharmacologic agents or mechanical prophylaxis be provided rather than no thromboprophylaxis. (3) The guidelines further recommend the use of pharmacologic prophylaxis during hospitalization, whether or not patients are using a compression device. The ACCP guidelines noted that compliance is a major issue with compression devices used for thromboprophylaxis and recommend that, if this prophylactic option is selected, use should be limited to portable, battery-operated devices. Moreover, it is recommended that devices be used for 18 hours per day. A 2009 non-randomized study found that there was better compliance with a portable battery-operated compression device compared to a non-mobile device when used by patients in the hospital following hip or knee replacement surgery. (4)

The ACCP also issued guidelines on VTE prophylaxis in non-orthopedic surgery patients. (5) For patients undergoing general or abdominal-pelvic surgery who have a risk of VTE of 3% or higher, the ACCP recommends prophylaxis with pharmacologic agents or intermittent compression rather than no prophylaxis. For patients at low risk for VTE (about 1.5%), the guidelines suggest mechanical prophylaxis. Unlike the guidelines on major orthopedic surgery, which recommends a minimum of 10-

14 days of VTE prophylaxis, the guideline on non-orthopedic surgery patients does not include a general timeframe for prophylaxis. They do, however, define “extended duration” pharmacologic prophylaxis as lasting 4 weeks; the latter is recommended only for patients at high risk for VTE, undergoing abdominal or pelvic surgery for cancer who are not otherwise at high risk for major bleeding complications.

National clinical guidelines have not specifically recommended use of compression devices in the outpatient setting. However, especially with the availability of portable, battery-operated devices, there is interest in use of outpatient compression devices for DVT following discharge from the hospital for major orthopedic and non-orthopedic surgery.

Regulatory Status

Multiple limb compression devices, with indications including prevention of DVT, have been cleared for marketing by the U.S. Food and Drug Administration (FDA) through the 510(k) process. Portable battery-operated devices that have been cleared by the FDA include:

Venowave VW5 (Venowave Inc.; Stouffville, Ontario, Canada): The device is a portable unit that is strapped to the leg below the knee. Different models are available that have a different maximum number of wave plate cycles per minute.

ActiveCare+SFT® System (Medical Compression Systems LTD, Or Akiva, Israel): The device applies sequential compression to the lower limb; it has the option of being battery-operated. Foot compression is achieved with use of a single-celled foot sleeve. Calf and thigh compression requires use of a 3-celled cuff sleeve.

Restep DVT System (Stortford Medical LLC, West Windsor, NJ): This is a lightweight device that utilizes single chamber pressure cuffs attached to the patient’s lower legs.

Kendall SCD™ 700 Sequential Compression System (Covidien, Mansfield, MA): This pneumatic compression device can be used in the clinic or at home. It has a 2-pronged plug and is not battery-operated.

Rationale

The policy was created in November 2012 with a search of the MEDLINE database and updated with a literature search through November 18, 2013. The key published literature is summarized below.

Venous thromboembolism prophylaxis in major orthopedic surgery patients

Patients without a contraindication to prophylaxis with pharmaceutical agents

Anticoagulation is the mainstay of deep vein thrombosis (DVT) prophylaxis after major surgery and is sometimes continued into the outpatient setting. Treatment with limb compression devices may offer additional benefit when used in conjunction with anticoagulation in the inpatient setting but is not commonly used in the outpatient setting. The ideal study design to evaluate whether there is benefit in the outpatient setting would be a randomized controlled trial (RCT) comparing outpatient anticoagulation alone to anticoagulation plus a limb compression device. Key health outcomes include incidence of DVT and pulmonary embolism (PE), as well as measures of functional status and/or quality of life associated with these outcomes.

Randomized controlled trials: No randomized controlled trials (RCTs) with the above design were identified. In 2012, Kakkos et al published a meta-analysis of RCTs evaluating combined use of anticoagulation and mechanical DVT prophylaxis following joint replacement surgery; however, the study focused on inpatient thromboprophylaxis.(7) The authors identified 4 trials that compared anticoagulation alone to anticoagulation plus use of limb compression devices. Three of the 4 studies used limb compression devices only until discharge from the hospital. In the fourth study, the article did not clearly state that that pump use was limited to the inpatient setting, but inpatient use was implied,

eg, the article stated that staff checked several times a day to ensure correct use of the pump system. Meta-analyses found statistically significantly lower incidences of DVT in the group that used compression pumps in addition to anticoagulation compared to anticoagulation-only. In a pooled analysis of 4 trials on hip replacement, the incidence of DVT was 9.7% in the anticoagulation-only group and 0.9% in the combined treatment group (risk ratio [RR], 0.17; 95% confidence interval [CI], 0.06 to 0.46). Similarly, when findings from 2 trials on knee replacement were pooled, the incidence of DVT was 18.7% in the anticoagulation-only group and 3.7% in the combined treatment group (RR=0.27; 95% CI, 0.08 to 0.89).

A 2013 systematic review and meta-analysis by Sobieraj et al was similar to the Kakkos et al study, described above. It reviewed studies comparing combined pharmacologic and mechanical prophylaxis to either method alone in patients undergoing major orthopedic surgery.(8) Six RCTs were identified and these studies all focused on inpatient treatment. Most studies evaluated outcomes in the postoperative period. In 1 study that followed patients for 90 days, intermittent limb compression continued only until hospital discharge.

There are several reasons why the benefit of limb compression devices in the hospital setting may not extrapolate to benefit in the outpatient setting. First, the level of mobility is necessarily less in the hospital than in the outpatient setting, indicating a different risk for DVT. Also, the use of limb compression devices in the hospital can be more highly controlled and monitored. In the outpatient setting, there are questions about the degree of compliance with the devices, including the ability to correctly use them in the absence of professional supervision. No comparative studies were identified that focused on compliance with limb compression devices in the outpatient setting.

Case series: A 2006 case series by Giannoni et al in Italy included both inpatient and outpatient DVT prophylaxis with limb compression devices and anticoagulants.(9) The study included 34 patients who underwent total knee replacement (4 patients had bilateral replacements). All patients used a compression device (A-V Impulse foot pump system) for 15 days. The mean hospital stay was 7 days, and the range was 5 to 12 days. The compression devices were worn for an average of 14 hours per day (range, 8-18 hours). Patients were also treated with low molecular-weight heparin (LMWH), beginning after surgery and continuing until the operated leg was completely weight bearing (15-30 days). Ultrasonography detected DVTs in 3 of 34 (8.8%) patients; all were distal DVTs. One symptomatic DVT developed on the 4th postoperative day, and there were 2 subclinical DVTs detected at the routine 1 month ultrasonographic examination. This study did not include a comparison group of patients who did not use a limb compression device. In addition, given the range of length of hospital stay, some patients received their entire course of prophylactic treatment as inpatients. Compliance with compression devices was not reported.

Patients with a contraindication to prophylaxis with pharmaceutical agents

Patients with contraindications to anticoagulants need to be treated with nonpharmacologic measures. The ideal study design for this question would be an RCT comparing prophylaxis with limb compression devices alone in the outpatient setting to no prophylaxis or to alternative methods of prophylaxis.

Randomized controlled trials: No RCTs using this design were identified. However, 1 recent RCT provided data that might be useful for answering the question of whether outpatient use of limb compression devices are beneficial in the absence of outpatient anticoagulant use. The study, reported on in 2 publications, one in 2010 and the other in 2011, was conducted at multiple centers in the United States and included 395 patients undergoing total hip replacement.(10,11) Individuals with a previous history of thrombosis, known coagulation disorder, solid malignant tumor, peptic ulcer disease or mental disorder were excluded. Patients were randomized to 10 days of DVT prophylaxis using either LMWH or a mobile limb compression device (ActiveCare+SFT). Treatment continued until 10 days after surgery in both groups; patients received a variable portion of their treatment after hospital discharge. Patients in the compression device group could also receive aspirin if recommended by their doctor. Patients were examined with bilateral duplex ultrasound on day 10 to 12 following surgery. The mean length of hospital stay was 3.2 days in both groups. Length of hospital stay ranged from 2 days to 10

days; thus, patients had between 0 days and 8 days of outpatient use of their assigned method of prophylaxis. According to ultrasound findings, 8 of 196 (4.1%) in the limb compression group and 8 of 190 (4.2%) in the LMWH group had a DVT. In addition, 2 pulmonary emboli were detected in each group. The incidence of venous thromboembolic events did not differ significantly between groups. However, the rate of major bleeding was significantly higher in the LMWH group. A total of 11 (6%) of patients in the LMWH group had a major bleeding event compared to no patients in the limb compression group ($p < 0.001$). Rates of minor bleeding were similar in the 2 groups; 78 (40%) in the LMWH group and 74 (37%) in the limb compression group. In addition, compliance with the mobile compression devices was monitored using internal timers in the device. According to these data, patients used the device for a mean of 11 days (range, 1-15 days) and for a mean of 20 hours per day. Mean use of the device was 83% of possible usable time. Findings on compliance were not reported separately for inpatient and outpatient use of the devices.

Section Summary: There is very little published evidence on the efficacy of outpatient use of limb compression devices for DVT prophylaxis after major orthopedic surgery. There are no RCTs that evaluate outpatient use of limb compression as an adjunct to pharmacologic prophylaxis in patients without a contraindication to anticoagulants. Some RCTs have evaluated the inpatient use of limb compression as an adjunct to pharmacologic agents, but the results of these trials might not be able to be extrapolated to the outpatient setting. There is also a lack of evidence on compliance with limb compression devices in the outpatient setting. National clinical guidelines support the use of limb compression devices DVT prophylaxis after major orthopedic surgery in patients who are not candidates for pharmacologic prophylaxis due to a high risk of bleeding. In addition, 1 RCT that reported similar rates of postoperative DVT in patients who received limb compression devices or LMWH provides some evidence in support of limb compression devices as the sole intervention in the outpatient setting. This study was limited in that much of the treatment occurred in the hospital, and patients with a known coagulation disorder were excluded from participation.

Venous thromboembolism prophylaxis in major nonorthopedic surgery patients

Patients with and without a contraindication to prophylaxis with pharmaceutical agents

Randomized controlled trials: No RCTs were identified that specifically addressed the comparison between inpatient-only and inpatient and outpatient use of limb compression devices as an adjunct to anticoagulant use in patients undergoing major nonorthopedic surgery. Moreover, no RCTs were identified that compared intermittent pneumatic compression (IPC) in the outpatient setting to no prophylaxis beyond inpatient use in patients with contraindications to pharmaceutical agents. Two systematic reviews of RCTs on venous thromboembolism (VTE) prophylaxis in patients undergoing major nonorthopedic surgery were examined, one a Cochrane review on VTE prevention in high-risk patients and the other on VTE prevention after gynecologic surgery; neither meta-analysis included RCTs relevant to the research question being considered.(12,13)

However, an RCT by Sobieraj-Teague et al may contribute some relevant data.(14) The nonblinded study, conducted in Canada, compared inpatient and outpatient use of Venowave, a portable battery-operated IPC device, to usual care only in 150 adult patients undergoing cranial or spinal neurosurgery. As part of usual care, all patients were prescribed graduated compression stockings and early mobilization. Patients could also receive pharmacologic treatment at the discretion of their physician. A total of 19 of 75 patients (25%) in the Venowave group and 26 of 75 patients (35%) in the control group received anticoagulants (unfractionated or LMW heparin) and an additional 4 (5%) in the Venogram group and 7 (9%) in the control group used aspirin. In the Venowave group, devices were worn until development of VTE, patient refusal, until undergoing a screening bilateral venogram at day 9 (± 2 days) or earlier if patients were discharged from the hospital earlier and were unwilling to return for a venogram. The median day of hospital discharge was day 4. Patients who continued using the Venowave device at home received home visits at least daily to optimize compliance. Eight patients did not undergo screening venography. Mean time to screening was 7.3 days in the Venowave group and 7.5 days in the control group. The primary efficacy outcome was a composite of asymptomatic DVTs

and symptomatic PEs. VTE occurred in 3 patients (4%) in the Venowave group and 14 (19%) in the control group. The difference between groups was statistically significant (RR=0.21; 95% CI, 0.05 to 0.75). Most of the VTEs were asymptomatic, and there were no PEs. Two patients in the control group and none in the Venowave group experienced a symptomatic DVT. Among the 75 evaluable patients in the Venogram group, 17 (23.3%) were continuous users, 39 (53.4%) were intermittent users, and 17 (23.3%) discontinued use of the device before their venogram assessment. Compliance might have been lower if patients had not received daily home visits.

The Sobieraj-Teague study did not specifically exclude patients with a contraindication to pharmaceutical agents. Moreover, only about 30% of participants were prescribed heparin or aspirin. This suggests that study findings might be applicable to patients who are not taking pharmaceutical agents, ie, including those with a contraindication. Generalizability of study findings is not clear, however, as the authors did not report VTE prevalence among patients who did or did not take anticoagulants or aspirin.

Section Summary: There is very little published evidence on the efficacy of outpatient use of limb compression devices for DVT prophylaxis after major nonorthopedic surgery. There are no RCTs that evaluate outpatient use of limb compression as an adjunct to pharmacologic prophylaxis in patients without a contraindication to anticoagulants. There is also a lack of evidence on compliance with limb compression devices in the outpatient setting. National clinical guidelines support the use of limb compression devices for DVT prophylaxis after major nonorthopedic surgery in individuals at moderate- and high risk of DVT. Moreover, 1 RCT, which found significantly fewer VTEs in patients undergoing cranial or neurosurgery who used a portable IPC than those receiving usual care; treatment occurred in both the inpatient and outpatient settings.

Summary

Patients undergoing major surgery, particularly orthopedic surgery, are at high risk for venous thromboembolism (VTE), and VTE prophylaxis for high-risk patients may be indicated beyond the period of hospitalization. Pharmacologic prophylaxis is the mainstay of treatment, but some patients have contraindications to anticoagulation, such as a high bleeding risk. For these patients who are undergoing major orthopedic surgery or other high-risk surgeries, limb compression devices are a reasonable alternative when prophylaxis is indicated in the outpatient setting. This is based on support in clinical practice guidelines, evidence from RCTs on different populations, and the lack of other good alternatives. Therefore, the use of limb compression devices for outpatient VTE prophylaxis may be considered medically necessary when prophylaxis is indicated but there are contraindications to anticoagulation.

For patients who do not have contraindications to anticoagulation, the evidence is not sufficient to determine whether limb compression devices offer additional benefit. There is a lack of studies that evaluate the incremental benefit of limb compression devices in addition to anticoagulants and a lack of evidence on outpatient compliance. Therefore, outpatient use of limb compression devices for VTE prophylaxis after major orthopedic surgery in patients who do not have a contraindication to pharmacologic prophylaxis is considered investigational.

Practice Guidelines and Position Statements

In 2012, the American College of Chest Physicians (ACCP) published updated evidence-based guidelines on prevention of VTE in orthopedic surgery and nonorthopedic surgical patients.

ACCP recommendations on use of limb compression devices in orthopedic surgical patients (3):

- "2.1.1 In patients undergoing total hip arthroplasty (THA) or total knee arthroplasty (TKA), we recommend use of one of the following for a minimum of 10 to 14 days rather than no antithrombotic prophylaxis: low-molecular-weight heparin (LMWH), fondaparinux, apixaban, dabigatran, rivaroxaban,

low-dose unfractionated heparin (LDUH), adjusted-dose vitamin K antagonist(VKA), aspirin (all Grade 1B) , or an intermittent pneumatic compression device (IPCD) (Grade 1C).”

- “2.1.2 In patients undergoing hip fracture surgery (HFS), we recommend use of one of the following rather than no antithrombotic prophylaxis for a minimum of 10 to 14 days: LMWH, fondaparinux, LDUH, adjusted-dose VKA, aspirin (all Grade 1B) , or an IPCD (Grade 1C).”

- “2.5. In patients undergoing major orthopedic surgery, we suggest using dual prophylaxis with an antithrombotic agent and an IPCD during the hospital stay (Grade 2C).

- “2.6. In patients undergoing major orthopedic surgery and increased risk of bleeding, we suggest using an IPCD or no prophylaxis rather than pharmacologic treatment (Grade 2C).”

For all of the above recommendations related to pneumatic compression pumps, the ACCP recommended only portable, battery-powered devices be used and stated that efforts should be made to ensure devices are worn for 18 hours per day. The authors noted that compliance is the biggest challenge associated with use of pneumatic compression devices.

ACCP recommendations on use of limb compression devices in nonorthopedic general and abdominal-pelvic surgical patients, stratified by patient risk of VTE and risk of bleeding (5):

Note: A recommended standard duration of prophylaxis was not defined. However, “extended duration” prophylaxis was defined as lasting 4 weeks.

Very low risk patients (<0.5%): “We recommend that no specific pharmacologic (Grade 1B) or mechanical (Grade 2C) prophylaxis be used other than early ambulation.”

Low risk for VTE (about 1.5%): “We suggest mechanical prophylaxis, preferably with intermittent pneumatic compression (IPC), over no prophylaxis (Grade 2C).”

Moderate risk for VTE (about 3%) and not at high risk of bleeding: “We suggest low-molecular-weight heparin (LMWH) (Grade 2B), low-dose unfractionated heparin (LDUH) (Grade 2B), or mechanical prophylaxis, preferably with IPC (Grade 2C), over no prophylaxis.”

Moderate risk for VTE (about 3%) and high risk for major bleeding complications or in whom bleeding consequences would be particularly severe: “We suggest mechanical prophylaxis, preferably with IPC, over no prophylaxis (Grade 2C).”

High risk for VTE (about 6.0%) and not at high risk of bleeding: “We recommend pharmacologic prophylaxis with LMWH (Grade 1B) or LDUH (Grade 1B) over no prophylaxis. We suggest that mechanical prophylaxis with elastic stockings (ES) or IPC should be added to pharmacologic prophylaxis (Grade 2C).”

High risk for VTE (about 6.0%) and high risk for major bleeding complications or in whom bleeding consequences would be particularly severe: “We suggest use of mechanical prophylaxis, preferably with IPC, over no prophylaxis until the risk of bleeding diminishes and pharmacologic prophylaxis may be initiated (Grade 2C).”

High- risk for VTE, both LMWH and unfractionated heparin contraindicated or unavailable and not at high risk for major bleeding complications: “We suggest low-dose aspirin (Grade 2C), fondaparinux (Grade 2C), or mechanical prophylaxis, preferably with IPC (Grade 2C), over no prophylaxis.”

High-risk for VTE, undergoing abdominal or pelvic surgery for cancer and not otherwise at high risk for major bleeding complications: “We recommend extended duration pharmacologic prophylaxis (4 weeks) with LMWH over limited-duration prophylaxis (Grade 1B).”

In 2011, the American Academy of Orthopaedic Surgeons (AAOS) published an updated guideline on prevention of venous thromboembolism in patients undergoing elective hip and knee arthroplasty. (6) The guideline included the following recommendations relevant to this policy:

“- The work group suggests the use of pharmacologic agents and/or mechanical compressive devices for the prevention of venous thromboembolism in patients undergoing elective hip or knee arthroplasty, and who are not at elevated risk beyond that of the surgery itself for venous thromboembolism or bleeding. (Grade of Recommendation: Moderate)

- Current evidence is unclear about which prophylactic strategy (or strategies) is/are optimal or suboptimal. Therefore, the work group is unable to recommend for or against specific prophylactics in these patients. (Grade of Recommendation: Inconclusive)

- In the absence of reliable evidence about how long to employ these prophylactic strategies, it is the opinion of this work group that patients and physicians discuss the duration of prophylaxis. (Grade of Recommendation: Consensus)

- In the absence of reliable evidence, it is the opinion of this work group that patients undergoing elective hip or knee arthroplasty, and who have also had a previous venous thromboembolism, receive pharmacologic prophylaxis and mechanical compressive devices. (Grade of Recommendation: Consensus)

- In the absence of reliable evidence, it is the opinion of this work group that patients undergoing elective hip or knee arthroplasty, and who also have a known bleeding disorder (eg, hemophilia) and/or active liver disease, use mechanical compressive devices for preventing venous thromboembolism. (Grade of Recommendation: Consensus)”

In 2007, the American College of Obstetricians-Gynecologists (ACOG) published a practice bulletin on prevention of DVT and PE after gynecologic surgery.(2) As with the ACCP recommendations described above, prophylaxis recommendations varied according to patient risk level. For patients at moderate and high risk of DVT, intermittent pneumatic compression was one of the recommended options for DVT prophylaxis. For patients at highest risk, ie, older than 60 years plus prior VTE, cancer or molecular hypercoagulable state, IPC or graduated compression stockings plus LDUH or LMWH was recommended as a prophylaxis option. For all but the highest risk patients, the practice bulletin stated that, when IPC devices were used, “the devices should be used continuously until ambulation and discontinued only at the time of hospital discharge.” For the highest risk patients, the document stated that continuing prophylaxis for 2 to 4 weeks after discharge should be considered.

Medicare National Coverage

There is no national coverage determination (NCD). In the absence of an NCD, coverage decisions are left to the discretion of local Medicare carriers.

References

1. Fisher WD. Impact of venous thromboembolism on clinical management and therapy after hip and knee arthroplasty. *Can J Surg* 2011; 54(5):344-51.
2. Committee on Practice Bulletins--American College of Obstetricians and Gynecologists. ACOG Practice Bulletin No. 84: Prevention of deep vein thrombosis and pulmonary embolism. *Obstet Gynecol* 2007; 110(2 Pt 1):429-40.
3. Falck-Ytter Y, Francis CW, Johanson NA et al. Prevention of VTE in orthopedic surgery patients: Antithrombotic Therapy and Prevention of Thrombosis, 9th ed: American College of Chest Physicians Evidence-Based Clinical Practice Guidelines. *Chest* 2012; 141(2 Suppl):e278S-325S.
4. Froimson MI, Murray TG, Fazekas AF. Venous thromboembolic disease reduction with a portable pneumatic compression device. *J Arthroplasty* 2009; 24(2):310-6.

5. Gould MK, Garcia DA, Wren SM et al. Prevention of VTE in nonorthopedic surgical patients: Antithrombotic Therapy and Prevention of Thrombosis, 9th ed: American College of Chest Physicians Evidence-Based Clinical Practice Guidelines. Chest 2012; 141(2 Suppl):e227S-77S.
6. American Academy of Orthopaedic Surgeons (AAOS). Preventing venous thromboembolic disease in patients undergoing elective hip and knee arthroplasty. Available online at: www.guideline.gov
7. Kakkos SK, Warwick D, Nicolaidis AN et al. Combined (mechanical and pharmacological) modalities for the prevention of venous thromboembolism in joint replacement surgery. J Bone Joint Surg Br 2012; 94(6):729-34.
8. Sobieraj DM, Coleman CI, Tongbram V et al. Comparative effectiveness of combined pharmacologic and mechanical thromboprophylaxis versus either method alone in major orthopedic surgery: a systematic review and meta-analysis. Pharmacotherapy 2013; 33(3):275-83.
9. Giannoni MF, Ciatti R, Capoccia L et al. Total knee replacement: prevention of deep-vein thrombosis using pharmacological (low-molecular-weight heparin) and mechanical (intermittent foot sole pump system) combined prophylaxis. Preliminary results. Int Angiol 2006; 25(3):316-21.
10. Colwell CW, Jr., Froimson MI, Mont MA et al. Thrombosis prevention after total hip arthroplasty: a prospective, randomized trial comparing a mobile compression device with low-molecular-weight heparin. J Bone Joint Surg Am 2010; 92(3):527-35.
11. Hardwick ME, Pulido PA, Colwell CW, Jr. A mobile compression device compared with low-molecular-weight heparin for prevention of venous thromboembolism in total hip arthroplasty. Orthop Nurs 2011; 30(5):312-6.
12. Kakkos SK, Caprini JA, Geroulakos G et al. Combined intermittent pneumatic leg compression and pharmacological prophylaxis for prevention of venous thromboembolism in high-risk patients. Cochrane Database Syst Rev 2008; (4):CD005258.
13. Rahn DD, Mamik MM, Sanses TV et al. Venous thromboembolism prophylaxis in gynecologic surgery: a systematic review. Obstet Gynecol 2011; 118(5):1111-25.
14. Sobieraj-Teague M, Hirsh J, Yip G et al. Randomized controlled trial of a new portable calf compression device (Venowave) for prevention of venous thrombosis in high-risk neurosurgical patients. J Thromb Haemost 2012; 10(2):229-35.

Billing Coding/Physician Documentation Information

E0650	Pneumatic compressor, nonsegmental home model
E0651	Pneumatic compressor, segmental home model without calibrated gradient pressure
E0652	Pneumatic compressor, segmental home model with calibrated gradient pressure
E0655	Nonsegmental pneumatic appliance for use with pneumatic compressor, half arm
E0656	Segmental pneumatic appliance for use with pneumatic compressor, trunk
E0657	Segmental pneumatic appliance for use with pneumatic compressor, chest
E0660	Nonsegmental pneumatic appliance for use with pneumatic compressor, full leg
E0665	Nonsegmental pneumatic appliance for use with pneumatic compressor, full arm
E0666	Nonsegmental pneumatic appliance for use with pneumatic compressor, half leg
E0667	Segmental pneumatic appliance for use with pneumatic compressor, full leg
E0668	Segmental pneumatic appliance for use with pneumatic compressor, full arm
E0669	Segmental pneumatic appliance for use with pneumatic compressor, half leg
E0670	Segmental pneumatic appliance for use with pneumatic compressor, integrated, 2 full legs and trunk
E0671	Segmental gradient pressure pneumatic appliance, full leg
E0672	Segmental gradient pressure pneumatic appliance, full arm
E0673	Segmental gradient pressure pneumatic appliance, half leg
E0676	Intermittent limb compression device (includes all accessories), not otherwise specified

Additional Policy Key Words

N/A

Policy Implementation/Update Information

4/1/13 New policy; may be considered medically necessary

4/1/14 "Pneumatic" removed from policy statements and policy title. Major nonorthopedic surgery changed to "major nonorthopedic surgery or nonmajor orthopedic surgery" in 3rd and 4th policy statements. "Postsurgical" added to policy. E0675 removed from code table as it is for arterial insufficiency and this policy is specific to venous thromboembolism

State and Federal mandates and health plan contract language, including specific provisions/exclusions, take precedence over Medical Policy and must be considered first in determining eligibility for coverage. The medical policies contained herein are for informational purposes. The medical policies do not constitute medical advice or medical care. Treating health care providers are independent contractors and are neither employees nor agents Blue KC and are solely responsible for diagnosis, treatment and medical advice. No part of this publication may be reproduced, stored in a retrieval system or transmitted, in any form or by any means, electronic, photocopying, or otherwise, without permission from Blue KC.